

# CERTIFICATION

**SENSIRION**  
THE SENSOR COMPANY

Product: **SDP6xx Family**  
Description: **Differential Pressure Sensor**

The above mentioned products were calibrated to meet the specifications according to the corresponding Sensirion data sheet.  
Each device is individually calibrated and tested after calibration. Only devices which 100% pass the specifications defined in the data sheet are released.

As transfer standard for the SDP6xx calibration, Furness Controls type FC0510-200 and FC0510-2000 Microprocessor Micromanometers are used. These instruments are themselves subject to a scheduled calibration procedure and traceable to national standards.

Standards incorporated to establish this comparison are NIST traceable.



Date of Certification: **July 2009**

Valid till: **July 2010**

Zurich, July 2009

A handwritten signature in black ink, appearing to be the initials "d.g." followed by a long horizontal stroke.

\_\_\_\_\_  
CEO, Sensirion AG

A handwritten signature in black ink, consisting of several overlapping loops and a long horizontal stroke.

\_\_\_\_\_  
Quality Manager, Sensirion AG